

POROS 50 HS 1,000ML BULK PK

Product No. **1-3359-07**
 Lot No. **SAMPLE**
 Date of Manufacture **20OCT2016**

TEST	SPECIFICATION	RESULT
PHYSICAL INSPECTION		
Labels Correct and Intact		Pass
Caps Sealed		Pass
PRODUCT APPEARANCE		
White to Off-White Media Slurry		Pass
DYNAMIC BINDING CAPACITY SOP 100025401		
5% Breakthrough with Lysozyme	57.0 - 75.3 mg/mL	64.74 mg/mL
PROTEIN STANDARD SEPARATION SOP 100025403		
Chymotrypsinogen Conductivity	21.7 - 23.4 mS/cm	22.469 mS/cm
Cytochrome C Conductivity	29.3 - 31.9 mS/cm	30.867 mS/cm
Lysozyme Conductivity	38.9 - 43.2 mS/cm	41.654 mS/cm

Pharmaceutical Grade Reagent. For Manufacturing and Laboratory Use only.

Manufactured in compliance with our ISO 9001 and ISO 13485 certified quality management system. Site: 2 Preston Ct. Bedford, MA, USA

Life Technologies Corporation
 5781 Van Allen Way
 Carlsbad, CA, USA 92008
www.thermofisher.com
 For inquiries, contact us at cofarequests@thermofisher.com

**Quality Assurance
 Issued 20OCT16**

POROS 50 HS 1,000ML BULK PK

Product No. 1-3359-07
Lot No. SAMPLE
Date of Manufacture 20OCT2016

TEST	SPECIFICATION	RESULT
FTIR ANALYSIS SOP 3019		
Compares to reference spectrum		Pass
BIOBURDEN TEST SOP 4371392		
Total Bioburden	<= 100 CFU/mL	<5.24 CFU/mL
ENDOTOXIN TEST SOP BED3069		
Endotoxin	<= 1.0 EU/mL	<0.500 EU/mL

Pharmaceutical Grade Reagent. For Manufacturing and Laboratory Use only.

Manufactured in compliance with our ISO 9001 and ISO 13485 certified quality management system. Site: 2 Preston Ct. Bedford, MA, USA

Life Technologies Corporation
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com

**Quality Assurance
Issued 20OCT16**