

DECLARATION OF CONFORMITY

In respect to the following directives

Low Voltage Directive	2006/95/EC
EMC Directive	2004/108/EC
WEEE & RoHS Directive	2002/96/EC & 2002/95/EC

the manufacturer

ThermoFisher
SCIENTIFIC

Thermo CRS Ltd.
5344 John Lucas Drive
Burlington, Ontario Canada L7L 6A6
Phone: (905) 332-2000 Fax: (905) 332-1114

hereby declares that the product(s)

Product Name	Dimension4 Controller F01704	
Compatible Dim4 Peripherals	Thermotor Carousel	LPT
	VAL Mover	Flip Mover
	E-Stop Hub	Orbitor RS
EM Immunity	EN61326-1:2006 EN61000-4-2; EN61000-4-3; EN61000-4-4; EN61000-4-5; EN61000-4-6; EN61000-4-8; EN61000-4-11; EN61000-3-2; EN61000-3-3 Electromagnetic Compatibility Requirements – Electrical Equipment for Measurement, Control and Laboratory Use – Immunity Characteristics – Limits and Methods of Measurement	
	CISPR11:2009 + A1:2010; EN55011:2009 + A1:2010 Class A, Group I Industrial, Scientific, and Medical (ISM) Equipment	
EM Emissions	FCC Part 15, Subpart B, Class A – Unintentional Radiators	
Safety	EN61010-1:2004 CAN/CSA-C22.2 No. 61010-1-04 UL61010A-1	
	This equipment falls in category 9 of the WEEE directive, which is exempt from the RoHS directive.	
Notes	None	

Supplementary Information

Systems using Thermo CRS products should be evaluated for compliance with local standards for specific application compatibility.

The "Safe Use of the System" chapter in the user's guide provides information to protect the operator against injury. This must be consulted before using the product.

The MoverLink is designed for connecting Dim4 Peripherals to a PC via RS-232 or USB in laboratory applications. For applications where the MoverLink is controlling a mover that is handling volatile, bio-hazardous, or radioactive samples is required, the end-user must carry out a risk assessment to determine what further measures may be required to protect the operator from injury. Examples of other measures may include adding a fume hood, the use of Personal Protective Equipment (PPE – gloves, coats, breathing apparatus, etc.).

When and Where Issued:

April 25, 2012
Burlington, Ontario, Canada

**Contact established in the
Community authorized to
compile the technical file or
the relevant technical
documents**

Hansjoerg Haas

Hansjoerg Haas
Product Line Director, Site Leader
Laboratory Automation
Thermo Fisher Scientific

Thermo Fisher Scientific
Altrincham
1 St. George's Court
Altrincham, UK WA14 5TP

CE

12