

DECLARATION OF CONFORMITY

In respect to the following directives

Low Voltage Directive
EMC Directive
WEEE & RoHS Directive

2006/95/EC
2004/108/EC
2002/96/EC & 2002/95/EC

the manufacturer

ThermoFisher
SCIENTIFIC



Thermo CRS Ltd
Laboratory Automation
5250 Mainway
Burlington, Ontario
Canada L7L 5Z1
Phone: (905) 332-2000 Fax: (905) 332-1114

hereby declares that the product(s)

Product Name

T3 Carousel (Base Motor Module)	S02128
T3 Carousel, 5 Position Base Mount for Orbitor	F01960
T3 Carousel, 5 Position Spine Mount for Orbitor	F01962
T3 Carousel, 8 Position Base Mount for Orbitor	F01963
T3 Carousel, 8 Position Spine Mount for Orbitor	F01965

conform(s) to the following standards or other normative documents

EM Immunity

EN61326:1997+A1:1998+A2:2001
Electromagnetic Compatibility Requirements – Electrical Equipment for Measurement, Control and Laboratory Use – Immunity Characteristics – Limits and Methods of Measurement

EM Emissions

CISPR 11 / EN 55011, Class A, Group I
Industrial, Scientific, and Medical (ISM) Equipment

FCC

FCC Part 15, Subpart B, Class A – Unintentional Radiators

Safety

EN61010-1:2001
UL61010A-1, UL61010-1:2004

Environmental (RoHS)

This equipment, to the best of our knowledge, complies with European Directive 2002/95/EC on the Restriction of Hazardous Substances (RoHS).

China RoHS

Conforms to standard GB/T26572. Refer to the following website for the information table.
<http://www.thermofisher.com/us/en/home/technical-resources/rohs-certificates.html>

Supplementary Information

Systems using Thermo CRS products should be evaluated for compliance with local standards for specific application compatibility.

The "Safe Use of the System" chapter in the user's guide provides information to protect the operator against injury. This must be consulted before using the product.

When and Where Issued:

June 27, 2016
Burlington, Ontario, Canada

Contact established in the Community authorized to compile the technical file or the relevant technical documents

Hansjoerg Haas

Hansjoerg Haas
Sr. Director Business Segment, Site Leader
Laboratory Automation
Thermo Fisher Scientific

Thermo Fisher Scientific Altrincham
1 St. George's Court
Altrincham, UK WA14 5TP

CE