

## CERTIFICATE OF ANALYSIS

### LINKS<sub>EQ</sub>™

**Product Name:** LinkS<sub>EQ</sub>™ HLA-B27 Typing Kit  
**Catalog Number:** 8011R-SL  
**Lot Number:** K4249-\*\*  
**Expiration Date:** See printed expiration date on package.


**Quality Control Requirement:** Each manufacturing lot of LinkS<sub>EQ</sub> product and reagents are tested against known samples. Each assay must yield correct reactions (positive or negative) for each sample tested.

**Method of Analysis:** LinkS<sub>EQ</sub> products use sequence-specific primers for the detection of alleles in human genomic DNA. Results were analyzed using SureTyper™ software.

**Results:** Passed

This product is prepared in accordance to GMP standards.

\*\*see kit for the suffixes

  
Christabel Aguele/Product Release Specialist

  
Date Issued

For Research Use Only-Not for Use in Diagnostic Procedures